



SUPPLIERS DECLARATION OF CONFORMITY

We, Lantronix, hereby declare that the product listed below, to which this Declaration of Conformity relates, is in conformity with the Standards and other Normative Documents listed below:

Product Type: MULTI-PORT MEDICAL DEVICE SERVER - EDS-MD
Product Number: EDSOR04P-01, EDSOR08P-01, EDSOR16P-01
Rated: 100-240 VAC, 50/60 Hz, 0.4A
Intended use: Commercial installations, indoor use

North America

Safety:

- ANSI/AAMI ES60601-1: 2005
- CAN/CSA-C22.2 No. 60601-1-08
- UL 60950-1-2011, 2nd Edition
- CAN/CSA C22.2 No. 60950-1-07, 2nd ed., 2006-07

Emissions:

- FCC Part 15, Subpart B, Class A
- ICES-003 Issue 6, Class A

European Union

Safety:

- EN 60601-1: 2006/A1: 2013
- IEC 60601-1: 1988 + A1: 1991 + A2: 1995

Emissions: EMC Directive (2014/30/EU)

- EN 60601-1-2: 2007
- EN 61000-3-2: 2014
- EN 61000-3-3: 2013
- EN 55032: 2012/AC: 2013, Class A

Immunity: EMC Directive (2014/30/EU)

- EN 60601-1-2: 2007
- EN 55024: 2010
- EN 61000-4-2: 2009
- EN 61000-4-3: 2006 + A1: 2008 + A2: 2010
- EN 61000-4-4: 2004
- EN 61000-4-5: 2006
- EN 61000-4-6: 2009
- EN 61000-4-8: 2010
- EN 61000-4-11: 2004

Other Countries

• Australia

- CISPR 11: 2003 + A1: 2004 + A2: 2006
- CISPR 32: 2012, Class A Emissions
- AS/NZS CISPR 22: 2009, Class A Emissions

• Japan

- VCCI V-3: 2010-04, Class A Emissions
- National Deviation to IEC 60601-1: 1988 + A1: 1991 + A2: 1995 based on JIST 0601-1

"Lantronix, 7535 Irvine Center Drive, Suite 100, Irvine, CA 92618, USA declares that the equipment specified above conforms to the referenced EU Directives and Harmonized Standards."

Signature: Daryl R. Miller Date: 8-17-18

Name: Daryl R. Miller Title: VP of Engineering