



EDS-MD[®] Medical Device Server Quick Start Guide

WHAT'S IN THE BOX



Quick Start Guide



EDS-MD[®] Medical Device Server with Ethernet Port Dust Covers




RJ45 CAT5E Ethernet Cable



RJ45 Loopback Adapter



Mounting Tabs (2)



Rubber feet (4)

1 INSTALL LANTRONIX PROVISIONING MANAGER

1. Download the latest version of Lantronix Provisioning Manager from <https://www.lantronix.com/products/lantronix-provisioning-manager/>
2. Install Lantronix Provisioning Manager.

For detailed installation instructions, see the Lantronix Provisioning Manager online help at <https://docs.lantronix.com/products/lpm/>


2 QUICK CONNECT

If your PC and EDS-MD can be connected to a network with a DHCP server:

1. Connect an Ethernet cable to a port on the switch and to Ethernet Main 1 on the EDS-MD.
2. Power up the unit with the appropriate regional power cord.*

If you are unable to connect the PC and EDS-MD to a network:

1. Use an Ethernet cable to connect the computer directly to Ethernet Main 1 on the EDS-MD.
2. Power up the unit with the appropriate regional power cord.*



* Power cords sold separately

3 CONFIGURATION

1. Run Lantronix Provisioning Manager.
2. If this is your first time running Lantronix Provisioning Manager, proceed through the initial setup.
3. Locate your device in the device list and note the device IP address.
4. Access Web Manager by entering the device IP address in the address bar of a browser. The device's factory default username is **"admin"** and factory default password is the last 8 characters of the Device ID (for devices manufactured after January 1, 2020) or **"PASS"** (for all older devices).
5. Assign an IP address in the network settings and perform other configuration.

Please see the EDS-MD product User Guide and Command Reference for detailed instructions: <http://www.lantronix.com/support/downloads>

If Lantronix Provisioning Manager is unable to discover the EDS-MD:

- Check that the "link" light is on when the RJ45 cable is connected to the computer and EDS-MD
- In Windows Device Manager, ensure that the Ethernet adapter for the local connection is installed and working properly

CONTACT

Technical Support

For technical support queries, visit <http://www.lantronix.com/support> or call (800) 422-7044 Monday - Friday from 6:00 a.m. - 5:00 p.m., Pacific Time, excluding holidays.

Latest Firmware

For the latest firmware downloads, visit <http://www.lantronix.com/support/downloads>

IP Address Assignment Tutorial

To view a tutorial on how to assign an IP address, visit <http://www.lantronix.com/support/tutorials>